



## EC – Declaration of Conformity

<b>Manufacturers Name:</b>	Orfit Industries N.V.
<b>SRN (Single Registration Number):</b>	BE-MF-000007872
<b>Manufacturers Address:</b>	Vosveld 9a, 2110 Wijnegem, Belgium
<b>Basic UDI-DI:</b>	54200287002145
<b>Name of the Device(s):</b>	Raycast® High Precision - Efficast® Masks, Nanor Masks and Hybrid Masks with Nanor for Patient Immobilisation -Head, Neck and Shoulder
<b>Product code(s):</b>	35763/16MI, 35763/2MI, 35763/16MI/NH, 35763/16MI/EM, 35760/EFF16MI, 35760/EFF16MI/EM, 35779/16MI, 35763/16MI+N, 35763/2MA, 35763/2MA/NH, 35763/2MA/M, 35763/2MA/EM, 35760/EFF2MA, 35779/2MA, 35779/2MA/M, 35779/2MA/EM, 35764/16MI, 35764/2MA, 35764/2MA/NH, 35764/2MA/M, 35764/2MI+N, 33700/16MI, 33700/16MI/M, 33700/16MI/EM, 33700/2MA, 33700/2MA/M, 33700/2MA/EM, 33700/2MA/NH, 33730/4, 33700/2MI+N, 33710/2MA, 33688/2MA, 33705/2MA/NH, 33688/2MA, 33705/2MA, 33702/2MA, 33794/2MA, 33759/16MI/12MI+N, 33740/2MA/12MI+N, 33740/2MA/12MI+N/NH, 33737/2MA/12MI+N/NH, 33733/16MI/12MI+N, 33730/2MA/12MI+N, 33730/2MA/12MI+N/NH, 33730/2MA/12MI+N/G17, 33730/2MA/12MI+N/NH/G17, 33791/2MA/12MI+N, 33782/16MI/12MI+N, 33783/16MI/12MI+N, 33748/2MA/12MI+N/NH, 33749/2MA/12MI+N/NH, 33785/16MI/12MI+N, 33787/16MI/12MI+N, 33750/2MA/12MI+N/NH, 33776/2MA, 33776/32MA, 35763/16MI+N/NH, 33700/2MI+N/NH
<b>Classification:</b>	Class I, according the rules of Annex VIII
<b>Conformity assessment route:</b>	Orfit Industries N.V. uses the following procedures for the CE-labelling of their products according the Regulation MDR 2017/745:  Class I: EC conformity declaration according to Annex IV.
<b>Applied norms:</b>	ISO 13485:2016 ISO 14971:2019 ISO 15223-1:2021 ISO 10993-5:2009 ISO 10993-10:2010

This declaration of conformity is issued under the sole responsibility of Orfit Industries N.V. We hereby declare that the medical device(s) specified above meet the provisions of the Regulation (EU) MDR 2017/745 for medical devices. This declaration is supported by the Quality System approval ISO 13485:2016 issued by Lloyd's Register EMEA – LRQA. All supporting documentation is retained at the premises of the manufacturer.

Eddy Marivoet,  
Wijnegem, 5 November 2021

Quality Assurance & Regulatory Affairs Manager

